

NordVal Validation

NordVal
c/o Department of Microbial Food Safety
Danish Institute for Food and Veterinary Research
Mørkhøj Bygade 19
DK-2860 Søborg
Denmark

Ref. No. : 2006-30-5408-00059

1st Approval: 28th April 2004

Renewal date: 28th April 2006

Valid until: 28th April 2008

Renewal of NordVal Certificate for Bioline Salmonella Elisa Test Selecta.

Bioline ApS, Fredericiavej 414, 7080 Børkop, Denmark has applied for renewal of the NordVal certificate for Bioline Salmonella Elisa Test Selecta for detection of Salmonella in all food matrices and animal feed. The Bioline Salmonella Elisa Test Selecta is produced by Bioline ApS, Denmark.

NordVal has studied the enclosures to the application and evaluated the results obtained in the validation by AFNOR (the French system for validation of alternative microbiological method) . ISO 16140 was used for the comparative part of the study in the AFNOR validation.

NordVal has concluded that it has been satisfactorily demonstrated that the requirements of the NordVal validation protocol are fulfilled for foods and animal feed. The results document the equal performance of Bioline Salmonella Elisa Test Selecta and the reference method: ISO 6579:2002.

In the certificate to the applicant it is stated, that NordVal should be informed of any changes of the method, and that every two years an application for renewal of the approval should be forwarded to NordVal. For the renewal of the NordVal certificate the method must be validated against the latest version of the reference method.

On behalf of the NordVal Steering Group


Sven Qvist

Chairman

Date: 26/4-2006