

Detection of non-salmonella strains - Bioline Salmonella SELECTA

Data were obtained by AFNOR expert laboratory validation. 30 non-salmonella strains were enriched and tested with the Bioline Salmonella test ELISA method

N°	Strain	OD	Result
1	<i>Bacillus cereus</i>	0.010	-
2	<i>Citrobacter diversus</i>	0.039	-
3	<i>Citrobacter freundii</i>	0.028	-
4	<i>Citrobacter freundii</i>	0.032	-
5	<i>Citrobacter freundii</i>	0.089	-
6	<i>Citrobacter freundii</i>	0.028	-
7	<i>Citrobacter youngae</i>	0.013	-
8	<i>Enterobacter agglomerans</i>	0.026	-
9	<i>Enterobacter cloacae</i>	0.022	-
10	<i>Enterobacter cloacae</i>	0.022	-
11	<i>Enterobacter cloacae</i>	0.033	-
12	<i>Enterobacter cloacae</i>	0.024	-
13	<i>Enterobacter cloacae</i>	0.031	-
14	<i>Enterobacter sakazakii</i>	0.025	-
15	<i>Escherichia coli</i>	0.039	-

N°	Strain	OD	Result
16	<i>Escherichia coli</i>	0.018	-
17	<i>Escherichia coli</i>	0.037	-
18	<i>Escherichia coli</i>	0.006	-
19	<i>Escherichia coli</i> O157:H7	0.013	-
20	<i>Hafnia alvei</i>	0.019	-
21	<i>Hafnia alvei</i>	0.007	-
22	<i>Hafnia alvei</i>	0.010	-
23	<i>Klebsiella oxytoca</i>	0.052	-
24	<i>Klebsiella pneumoniae</i>	0.015	-
25	<i>Proteus mirabilis</i>	0.018	-
26	<i>Proteus mirabilis</i>	0.004	-
27	<i>Providencia rettgeri</i>	0.023	-
28	<i>Pseudomonas aeruginosa</i>	0.032	-
29	<i>Shigella sonnei</i>	0.004	-
30	<i>Yersinia enterocolitica</i>	0.000	-